


ENDODONTIC SURGERY: FILLING OF PERIAPICAL CAVITIES

Specific indications and instructions for Guided Regeneration Material System


COLLAGENE AT® IDROSSILAPATITE AT® IDROSSILAPATITE MICROMETRICA AT®

The use of guided regeneration materials Idrossilapatite AT® and Collagene AT® allows new bone growth in the periapical zone without residual cavity.


1. After apicectomy is performed, remove all soft tissues inside the periapical lesion.



2. Pass the sutures without tying and prepare the guided regeneration AT materials.




3. Put Idrossilapatite AT® in to the cavity and press with flat obturators and dry gauzes till the complete filling of the periapical cavity.




ENDODONTIC SURGERY: FILLING OF PERIAPICAL CAVITIES

4. Apply a Collagene AT® membrane of suitable sizes, covering the bone margins for about two millimeters.



5. Adapt the membrane with blunt dry instruments and dry gauzes.



6. Tie and complete the sutures so that the flap covers the membrane and the bone completely.



7. Prescribe a pharmacological treatment and advise to avoid chewing in the operated zone till sutures removal.

After one week remove sutures.

Warning: if wound dehiscence occurs, let the wound heal applying a surgical dressing for protection.

